



## 66 -- Millimeter Wave (MMW) Anechoic Chamber Refurbishment

- [Combine Synopsis/Solicitation](#) - Posted on Jun 20, 2006
- [Amendment to Combined Synopsis/Solicitation 01](#) - Posted on Jun 20, 2006

### General Information

Document Type: Amendment to Combined Synopsis/Solicitation  
Solicitation Number: N00173-06-R-CR04  
Posted Date: Jul 24, 2006  
Original Response Date: Jul 26, 2006  
Current Response Date: Jul 26, 2006  
Original Archive Date: Jun 20, 2007  
Current Archive Date: Jun 20, 2007  
Classification Code: 66 -- Instruments & laboratory equipment  
Set Aside: Total Small Business  
Naics Code: 334516 -- Analytical Laboratory Instrument Manufacturing

### Contracting Office Address

Department of the Navy, Office of Naval Research, Naval Research Laboratory, 4555 Overlook Ave. S.W., Washington, DC, 20375

### Description

The purpose of this amendment is to answer potential offerors questions.

Question 1: What is the shielding requirement for the pneumatic door that is to be refurbished;

100dB at 1GHz.

Answer to Question 1: Yes, that is the correct requirement, 100dB at 1GHz.

Question 2: Is a shielding effectiveness test required.

Answer to Question 2: Yes, the contractor shall verify shielding integrity/effectiveness once the door has been repaired.

Question 3: Is the shielding effectiveness test merely for the door area only or the entire chamber.

Answer to Question 3: The shielding effectiveness test is for the door area only. Once the door is repaired and inflated, the contractor shall probe all seams where covers were removed and probe the door area for leaks.

Question 4: Is baseline shielding effectiveness test required.

Answer to Question 4: A baseline test is not required.

### Point of Contact

Alan Crupi, Contract Specialist, Phone 202-767-3595, Fax 202-767-0430, Email wayne.crupi@nrl.navy.mil - F. Janilea Bays, Contracting Officer, Phone 202-767-2974, Fax 202-767-0430, Email jan.bays@nrl.navy.mil

### Place of Performance

Address: Naval Research Laboratory 4555 Overlook Ave SW Washington, DC  
Postal Code: 20375  
Country: UNITED STATES

[Register to Receive Notification](#)

---

### [Government-wide Numbered Notes](#)

#### **You may return to Business Opportunities at:**

- DON ONR listed by [\[Posted Date\]](#)
- DON Agencywide listed by [\[Posted Date\]](#)

---

[\[Home\]](#) [\[SEARCH synopses\]](#) [\[Procurement Reference Library\]](#)