



66 -- Versatile Laser Cutting System

- [Combine Synopsis/Solicitation](#) - Posted on Apr 19, 2007

General Information

Document Type: Amendment to Combined Synopsis/Solicitation
Solicitation Number: N00173-07-R-JK08
Posted Date: Apr 27, 2007
Original Response Date: May 21, 2007
Current Response Date: May 21, 2007
Original Archive Date: Apr 27, 2008
Current Archive Date: Apr 27, 2008
Classification Code: 66 -- Instruments & laboratory equipment
Naics Code: 334516 -- Analytical Laboratory Instrument Manufacturing

Contracting Office Address

Department of the Navy, Office of Naval Research, Naval Research Laboratory, 4555 Overlook Ave. S.W.,
Washington, DC, 20375, UNITED STATES

Description

The purpose of this amendment is to address potential offerors questions.

Question 1: Are there suppose to be deliverables annotated for CLIN0007?

Answer to Question 1: CLIN 0007 is for the extended warranty.

Question 1: Should the Specifications designate that the Nd:YAG is required for bruiting of the diamond instead of cutting as called out in the Specifications.

Answer to Question 1: The specification is for cutting, not bruiting. While I understand the issue with transparent samples such as diamond and glass, I am under the impression that transparent samples could be marked with a colored marker to unable cutting by a Nd:YAG laser.

2. Question 2: Can you please provide clarification of what is to be done to the glass with the Nd:YAG?

Answer to Question 2: Please see answer to Question 1

Question 3: Please provide clarification on item 12 referring to a saw cut. Is this supposed to refer to a laser cut instead of the saw cut?

Answer to Question 3: The statement requests a demonstration of the ability of the laser to cut through and across a silicon wafer from edge to edge as a 'cutting saw' would do.

Question 4: Is the solicitation unrestricted?

Answer to Question 4: Yes the solicitation is unrestricted.

Question 5: Would NRL consider making the solicitation a Small Business Set-aside?

Answer to Question 5: No the solicitation will remain unrestricted.

Original Point of Contact

Janiece Shall, Contract Specialist, Phone 202-767-3710, Fax 202-767-0430, Email janiece.shall@nrl.navy.mil - F.
Janilea Bays, Contracting Officer, Phone 202-767-2974, Fax 202-767-0430, Email jan.bays@nrl.navy.mil

Current Point of Contact

Janiece Shall, Contract Specialist, Phone 202-767-3710, Fax 202-767-0430, Email janiece.shall@nrl.navy.mil

Place of Performance

Address: Naval Research Laboratory 4555 Overlook Ave. SW Washington, DC 20375

Postal Code: 20375

Country: UNITED STATES

[Register to Receive Notification](#)

[Government-wide Numbered Notes](#)

You may return to Business Opportunities at:

- DON ONR listed by [\[Posted Date\]](#)
- DON Agencywide listed by [\[Posted Date\]](#)

[\[Home\]](#) [\[SEARCH synopses\]](#) [\[Procurement Reference Library\]](#)